
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001004
8:53 16/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:38
STERILIZING 3:49
136.4 C 232kPa 3:49
136.4 C 233kPa 4:19
136.3 C 232kPa 4:49
136.3 C 232kPa 5:19
136.3 C 232kPa 5:49
136.1 C 230kPa 6:19
136.3 C 232kPa 6:49
136.1 C 230kPa 7:19
136.1 C 230kPa 7:19
VENTING 7:19
AIR DRYING 8:09
CYCLE COMPLETE 8:18

Digital Signature #
5C100B74E25211BB
