

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3292

Name: RITTER 1  
Sterilizer ID: V2178198  
05/26/2023 03:02 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.5 F 0.0  
2:00 132.6 F 0.0  
4:00 153.8 F 0.2  
6:00 185.3 F 3.8  
8:00 199.8 F 8.2  
10:00 216.5 F 12.8  
12:00 225.2 F 7.2  
14:00 236.5 F 12.1  
16:00 235.6 F 9.2  
18:00 239.8 F 10.0  
20:00 247.1 F 13.6  
22:00 258.8 F 20.0  
24:00 269.3 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.1  
1:00 272.5 F 28.7  
2:00 271.4 F 28.0  
3:00 271.9 F 28.7  
4:00 272.5 F 28.7

Min 271.0 F 27.9  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.9 F 0.0  
5:00 206.8 F 0.0  
10:00 209.7 F 0.0  
15:00 214.3 F 0.0  
20:00 213.5 F 0.0  
25:00 216.3 F 0.0  
30:00 214.9 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 25:04  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:03:25  
CYCLE COMPLETE

05/26/2023 04:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_