

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 008119  
9:06 02/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:56  
PRESSURIZING 3:57  
STERILIZING 5:04  
136.3 C 221kPa 5:04  
136.3 C 223kPa 5:34  
136.0 C 217kPa 6:04  
136.4 C 222kPa 6:34  
136.4 C 221kPa 7:04  
136.0 C 219kPa 7:34  
136.4 C 222kPa 8:04  
136.3 C 221kPa 8:34  
136.3 C 221kPa 8:34  
VENTING 8:34  
AIR DRYING 9:25  
CYCLE COMPLETE 39:25

-----  
Digital Signature #  
F69595168EA1C1F6  
-----