
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000149
12:53 14/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:45
STERILIZING 5:22
136.0 C 220kPa 5:22
136.3 C 223kPa 5:52
136.1 C 223kPa 6:22
136.3 C 225kPa 6:52
136.0 C 223kPa 7:22
136.4 C 225kPa 7:52
136.4 C 225kPa 8:22
136.4 C 224kPa 8:52
136.4 C 224kPa 8:52
VENTING 8:52
AIR DRYING 9:53
CYCLE COMPLETE 10:32

Digital Signature #
FE14FDA00E14C7C4
