
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 001088
11:42 20/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:09
STERILIZING 4:18
136.0 C 222kPa 4:18
136.4 C 226kPa 4:48
136.3 C 224kPa 5:18
136.2 C 224kPa 5:48
136.3 C 224kPa 6:18
136.2 C 224kPa 6:48
136.2 C 224kPa 7:18
136.4 C 223kPa 7:48
136.4 C 223kPa 7:48
VENTING 7:48
AIR DRYING 8:43
CYCLE COMPLETE 23:43

Digital Signature #
3B2D40DD442F1359
