
Midmark M11 - v1.0.0
Total Cycles: 4004

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2023 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 71.9 F 0.1
4:00 78.6 F 0.8
6:00 101.7 F 2.9
8:00 134.8 F 5.8
10:00 162.3 F 9.1
12:00 183.7 F 12.9
14:00 203.1 F 3.4
16:00 215.1 F 6.7
18:00 225.2 F 10.2
20:00 233.0 F 9.2
22:00 236.9 F 10.2
24:00 237.1 F 8.8
26:00 246.6 F 13.5
28:00 257.6 F 19.5
30:00 267.8 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.1 F 29.0
2:00 271.7 F 28.3
3:00 272.2 F 29.1
4:00 271.9 F 28.5

Min 271.2 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.1 F 0.0
5:00 214.3 F 0.0
10:00 210.7 F 0.0
15:00 222.9 F 0.0
20:00 231.3 F 0.0
25:00 227.9 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:14
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:09:30
CYCLE COMPLETE

11/10/2023 09:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____