
Midmark M11 - v1.0.5
Total Cycles: 751

Name: RITTER2
Sterilizer ID: V2313228
02/02/2022 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.1 F 0.0
2:00 97.8 F 0.0
4:00 122.6 F 0.1
6:00 155.8 F 1.6
8:00 178.9 F 5.1
10:00 201.8 F 9.0
12:00 212.5 F 11.2
14:00 226.3 F 8.5
16:00 237.3 F 13.5
18:00 238.5 F 11.0
20:00 236.7 F 9.3
22:00 250.2 F 16.0
24:00 262.1 F 22.7
26:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.7 F 29.8
2:00 272.3 F 29.2
3:00 271.3 F 28.9
4:00 272.2 F 29.8

Min 271.3 F 28.7
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.8 F 0.0
5:00 210.6 F 0.0
10:00 217.6 F 0.0
15:00 214.9 F 0.0
20:00 207.8 F 0.0
25:00 213.3 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 26:27
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:04:25
CYCLE COMPLETE

02/02/2022 12:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____