
Midmark M11 - v1.0.5
Total Cycles: 3712

Name: RITTER2
Sterilizer ID: V2313228
06/18/2024 12:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.9 F 0.0
2:00 165.6 F 3.3
4:00 179.3 F 7.5
6:00 203.7 F 12.7
8:00 223.5 F 7.4
10:00 238.8 F 12.8
12:00 237.3 F 10.1
14:00 236.1 F 8.5
16:00 249.7 F 15.4
18:00 261.4 F 22.2
20:00 272.2 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.9
1:00 271.7 F 29.4
2:00 272.5 F 29.4
3:00 271.0 F 28.8
4:00 272.4 F 29.8

Min 271.0 F 28.7
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 234.1 F 0.0
10:00 220.9 F 0.0
15:00 216.0 F 0.0
20:00 216.0 F 0.0
25:00 223.0 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 00:58:60
CYCLE COMPLETE

06/18/2024 01:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____