

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3663

Name: RITTER 1  
Sterilizer ID: V2178198  
08/16/2023 11:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 167.8 F 0.0  
2:00 184.5 F 2.9  
4:00 177.3 F 6.8  
6:00 214.6 F 11.6  
8:00 221.9 F 5.6  
10:00 232.8 F 10.0  
12:00 237.1 F 8.5  
14:00 240.7 F 11.9  
16:00 239.0 F 10.0  
18:00 251.7 F 16.3  
20:00 262.8 F 22.7  
22:00 272.6 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 271.9 F 28.4  
2:00 272.2 F 28.7  
3:00 272.5 F 29.1  
4:00 271.8 F 28.8

Min 271.5 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.0 F 0.0  
5:00 223.1 F 0.0  
10:00 229.1 F 0.0  
15:00 227.1 F 0.0  
20:00 224.3 F 0.0  
25:00 223.2 F 0.0  
30:00 220.6 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 22:14  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:01:00  
CYCLE COMPLETE

08/16/2023 12:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_