
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002439
16:52 22/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:50
STERILIZING 3:51
136.0 C 221kPa 3:51
136.4 C 222kPa 4:21
136.4 C 226kPa 4:51
136.4 C 225kPa 5:21
136.3 C 224kPa 5:51
136.2 C 224kPa 6:21
136.0 C 223kPa 6:51
136.4 C 224kPa 7:21
136.4 C 224kPa 7:21
VENTING 7:21
AIR DRYING 8:12
CYCLE COMPLETE 15:35

Digital Signature #
168E66E2CADCF0F5
