
Midmark M11 - v1.0.0
Total Cycles: 929

Name: RITTER 1
Sterilizer ID: V2178198
06/23/2021 10:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.7 F 0.0
2:00 167.2 F 1.8
4:00 169.4 F 4.4
6:00 199.5 F 8.1
8:00 213.8 F 12.9
10:00 225.0 F 6.4
12:00 235.1 F 10.7
14:00 233.3 F 7.5
16:00 243.5 F 12.7
18:00 241.6 F 11.1
20:00 253.4 F 17.0
22:00 264.0 F 23.3
24:00 272.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.7 F 29.1
2:00 271.9 F 28.5
3:00 271.5 F 28.5
4:00 272.6 F 29.1

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.5 F 0.0
5:00 219.4 F 0.0
10:00 227.6 F 0.0
15:00 232.7 F 0.0
20:00 225.8 F 0.0
25:00 221.9 F 0.0
30:00 227.0 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 24:14
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:03:10
CYCLE COMPLETE

06/23/2021 11:48 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____