

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4290

Name: RITTER2  
Sterilizer ID: V2313228  
01/23/2025 05:38 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.1 F 0.0  
2:00 156.0 F 2.8  
4:00 180.9 F 6.9  
6:00 203.5 F 11.6  
8:00 218.6 F 5.6  
10:00 231.6 F 10.6  
12:00 234.0 F 8.0  
14:00 243.8 F 13.5  
16:00 243.5 F 12.5  
18:00 255.6 F 18.7  
20:00 266.1 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.2 F 29.3  
2:00 271.5 F 29.6  
3:00 271.9 F 29.2  
4:00 271.8 F 29.7

Min 271.1 F 28.9  
Max 272.3 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.7 F 0.0  
5:00 191.9 F 0.0  
10:00 194.4 F 0.0  
15:00 185.5 F 0.0  
20:00 182.7 F 0.0  
25:00 185.7 F 0.0  
30:00 187.9 F 0.0  
Drying Complete

FILLING: 1:41  
HEATING: 21:41  
STERILIZING: 4:00  
VENTING: 3:57  
DRYING: 30:00  
TOTAL CYCLE: 01:01:18  
CYCLE COMPLETE

01/23/2025 06:39 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_