
Midmark M11 - v1.0.0
Total Cycles: 1099

Name: RITTER 1
Sterilizer ID: V2178198
08/19/2021 10:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 180.5 F 0.0
2:00 184.1 F 3.3
4:00 208.0 F 8.3
6:00 223.4 F 8.9
8:00 238.0 F 9.8
10:00 233.6 F 8.1
12:00 245.5 F 12.7
14:00 245.4 F 12.7
16:00 258.6 F 19.6
18:00 269.9 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.0 F 28.2
2:00 272.6 F 28.7
3:00 271.3 F 27.9
4:00 272.2 F 28.5

Min 271.3 F 27.8
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.2 F 0.0
5:00 213.4 F 0.0
10:00 203.1 F 0.0
15:00 200.8 F 0.0
20:00 200.7 F 0.0
25:00 200.7 F 0.0
Dry Cycle Aborted

FILLING: 1:47
HEATING: 18:50
STERILIZING: 4:00
VENTING: 3:07
DRYING: 27:43
TOTAL CYCLE: 00:55:27

08/19/2021 11:45 AM
ACCEPTED: X
REJECTED:
Operator: U02