
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 006327
8:18 20/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:20
STERILIZING 4:23
136.4 C 222kPa 4:23
136.4 C 221kPa 4:53
136.1 C 222kPa 5:23
136.4 C 223kPa 5:53
136.4 C 223kPa 6:23
136.4 C 222kPa 6:53
136.4 C 222kPa 7:23
136.0 C 220kPa 7:53
136.0 C 220kPa 7:53
VENTING 7:53
AIR DRYING 8:40
CYCLE COMPLETE 8:55

Digital Signature #
5C0D47C3B65A4D69
