
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 003709
11:33 27/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:54
PRESSURIZING 1:57
STERILIZING 3:00
136.5 C 222kPa 3:00
136.2 C 224kPa 3:30
136.3 C 224kPa 4:00
136.2 C 221kPa 4:30
136.4 C 223kPa 5:00
136.4 C 221kPa 5:30
136.4 C 221kPa 6:00
136.2 C 221kPa 6:30
136.2 C 221kPa 6:30
VENTING 6:30
AIR DRYING 7:52
CYCLE COMPLETE 7:57

Digital Signature #
8183A68C8CDDEA4B
