
Midmark M11 - v1.0.5
Total Cycles: 108

Name: RITTER2
Sterilizer ID: V2313228
06/16/2021 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 110.6 F 0.0
2:00 111.9 F 0.9
4:00 119.4 F 3.1
6:00 145.7 F 6.3
8:00 173.1 F 10.3
10:00 197.2 F 6.1
12:00 210.5 F 7.5
14:00 224.0 F 12.0
16:00 231.3 F 8.8
18:00 241.9 F 11.6
20:00 242.7 F 12.8
22:00 254.7 F 18.8
24:00 265.0 F 25.2
26:00 270.7 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.9
1:00 271.5 F 29.9
2:00 271.6 F 29.9
3:00 271.8 F 29.8
4:00 271.9 F 30.0

Min 271.4 F 29.8
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.8 F 0.0
5:00 228.2 F 0.0
10:00 231.2 F 0.0
15:00 229.5 F 0.0
20:00 200.4 F 0.0
25:00 189.7 F 0.0
30:00 185.0 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 26:35
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:04:50
CYCLE COMPLETE

06/16/2021 02:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____