

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003544  
10:36 27/04/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:40  
STERILIZING 3:41  
136.4 C 224kPa 3:41  
136.3 C 223kPa 4:11  
136.4 C 223kPa 4:41  
136.4 C 225kPa 5:11  
136.1 C 217kPa 5:41  
136.2 C 219kPa 6:11  
136.4 C 224kPa 6:41  
136.1 C 222kPa 7:11  
136.1 C 222kPa 7:11  
VENTING 7:11  
AIR DRYING 8:02  
CYCLE COMPLETE 8:30  
-----

Digital Signature #  
1971EC1F93E691F8  
-----