
Midmark M11 - v1.0.5
Total Cycles: 238

Name: RITTER2
Sterilizer ID: V2313228
07/22/2021 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.9 F 0.0
2:00 164.4 F 3.4
4:00 187.5 F 7.9
6:00 209.3 F 13.4
8:00 223.3 F 7.6
10:00 237.0 F 13.0
12:00 238.9 F 11.1
14:00 238.7 F 10.1
16:00 252.6 F 17.0
18:00 264.4 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.5 F 29.2
2:00 272.1 F 29.5
3:00 272.3 F 29.1
4:00 272.0 F 29.4

Min 271.6 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.0 F 0.0
5:00 230.0 F 0.0
10:00 221.7 F 0.0
15:00 216.0 F 0.0
20:00 210.5 F 0.0
25:00 208.9 F 0.0
30:00 207.4 F 0.0
Drying Complete

FILLING: 2:07
HEATING: 19:56
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:13
CYCLE COMPLETE

07/22/2021 11:57 AM
ACCEPTED: X
REJECTED:
Operator: U02