

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2754

Name: RITTER2  
Sterilizer ID: V2313228  
08/11/2023 10:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.0 F 0.0  
2:00 70.6 F 0.0  
4:00 76.3 F 0.0  
6:00 107.4 F 0.0  
8:00 144.6 F 0.0  
10:00 180.2 F 0.8  
12:00 202.4 F 4.6  
14:00 214.9 F 8.5  
16:00 226.3 F 12.6  
18:00 230.0 F 8.2  
20:00 241.2 F 13.0  
22:00 240.6 F 10.9  
24:00 237.9 F 9.7  
26:00 252.2 F 16.6  
28:00 263.9 F 23.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.6  
1:00 271.6 F 28.7  
2:00 271.7 F 28.8  
3:00 271.9 F 29.0  
4:00 271.9 F 29.0

Min 271.6 F 28.5  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 234.7 F 0.0  
Dry Cycle Aborted

FILLING: 1:06  
HEATING: 29:54  
STERILIZING: 4:00  
VENTING: 2:46  
DRYING: 0:24  
TOTAL CYCLE: 00:38:09

08/11/2023 10:38 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_