

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1610

Name: RITTER2  
Sterilizer ID: V2313228  
10/17/2022 03:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
2:00 201.9 F 4.1  
4:00 207.1 F 9.7  
6:00 221.8 F 7.4  
8:00 233.8 F 11.8  
10:00 236.1 F 10.1  
12:00 236.1 F 9.0  
14:00 250.1 F 16.2  
16:00 262.2 F 23.2  
18:00 272.7 F 30.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.5  
1:00 272.5 F 30.2  
2:00 271.8 F 29.5  
3:00 271.5 F 29.9  
4:00 272.5 F 30.2  
  
Min 271.3 F 29.4  
Max 272.7 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.0 F 0.0  
5:00 211.3 F 0.0  
10:00 232.4 F 0.0  
15:00 233.8 F 0.0  
20:00 229.6 F 0.0  
25:00 224.6 F 0.0  
30:00 223.0 F 0.0  
Drying Complete

FILLING: 2:26  
HEATING: 18:21  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 00:58:06  
CYCLE COMPLETE

10/17/2022 04:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_