
Midmark M11 - v1.0.0
Total Cycles: 3333

Name: RITTER 1
Sterilizer ID: V2178198
06/05/2023 09:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.9 F 0.0
2:00 167.3 F 0.0
4:00 190.9 F 1.6
6:00 207.3 F 6.6
8:00 222.2 F 11.4
10:00 226.2 F 6.7
12:00 238.9 F 12.1
14:00 237.9 F 9.7
16:00 237.4 F 8.8
18:00 249.3 F 14.8
20:00 261.2 F 21.4
22:00 271.8 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.6 F 29.0
2:00 272.3 F 28.5
3:00 271.5 F 28.0
4:00 272.0 F 28.6

Min 271.4 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 208.6 F 0.0
10:00 227.5 F 0.0
15:00 228.3 F 0.0
20:00 221.6 F 0.0
25:00 218.1 F 0.0
30:00 215.8 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 22:39
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:45
CYCLE COMPLETE

06/05/2023 10:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____