
Midmark M11 - v1.0.0
Total Cycles: 2482

Name: RITTER 1
Sterilizer ID: V2178198
10/25/2022 06:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.9 F 0.0
2:00 86.8 F 0.0
4:00 95.2 F 0.1
6:00 131.9 F 2.0
8:00 157.1 F 4.8
10:00 181.3 F 8.0
12:00 201.6 F 11.5
14:00 213.3 F 2.9
16:00 222.7 F 6.7
18:00 232.2 F 10.4
20:00 233.3 F 8.4
22:00 240.1 F 11.1
24:00 235.6 F 8.4
26:00 249.3 F 14.7
28:00 260.2 F 20.8
30:00 270.2 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.5 F 28.7
2:00 271.3 F 28.3
3:00 272.0 F 28.6
4:00 271.4 F 28.5

Min 271.1 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
5:00 177.6 F 0.0
10:00 171.1 F 0.0
15:00 209.7 F 0.0
20:00 221.2 F 0.0
25:00 223.8 F 0.0
30:00 207.6 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:54
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:08:59
CYCLE COMPLETE

10/25/2022 08:07 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____