
Midmark M11 - v1.0.0
Total Cycles: 1852

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2022 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.3 F 0.0
2:00 117.5 F 0.1
4:00 133.1 F 2.2
6:00 161.3 F 5.4
8:00 186.6 F 9.4
10:00 214.6 F 7.9
12:00 224.0 F 6.8
14:00 233.9 F 11.4
16:00 231.8 F 7.3
18:00 244.4 F 12.9
20:00 241.7 F 10.9
22:00 254.1 F 17.2
24:00 265.0 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.4 F 28.5
2:00 271.6 F 28.3
3:00 272.5 F 28.6
4:00 271.5 F 28.1

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 226.2 F 0.0
10:00 222.3 F 0.0
15:00 224.0 F 0.0
20:00 213.4 F 0.0
25:00 214.3 F 0.0
30:00 222.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:53
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:03:58
CYCLE COMPLETE

04/21/2022 10:05 AM
ACCEPTED: X
REJECTED:
Operator: U02