
Midmark M11 - v1.0.5
Total Cycles: 2965

Name: RITTER2
Sterilizer ID: V2313228
10/04/2023 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.5 F 0.0
2:00 126.4 F 0.5
4:00 130.2 F 2.1
6:00 152.9 F 5.0
8:00 183.2 F 8.7
10:00 201.7 F 13.0
12:00 218.2 F 6.3
14:00 228.9 F 10.7
16:00 230.2 F 7.6
18:00 241.7 F 12.8
20:00 240.8 F 11.2
22:00 252.5 F 17.1
24:00 262.5 F 23.2
26:00 272.6 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.3 F 29.5
2:00 271.7 F 29.7
3:00 273.1 F 30.0
4:00 271.8 F 29.9

Min 270.9 F 29.0
Max 273.4 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.6 F 0.0
5:00 213.1 F 0.0
10:00 199.3 F 0.0
15:00 190.5 F 0.0
20:00 192.4 F 0.0
25:00 197.0 F 0.0
30:00 193.4 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 26:19
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:04:43
CYCLE COMPLETE

10/04/2023 11:03 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____