

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3381

Name: RITTER 1  
Sterilizer ID: V2178198  
06/13/2023 12:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.4 F 0.0  
2:00 163.0 F 0.0  
4:00 188.3 F 2.6  
6:00 213.8 F 7.6  
8:00 227.6 F 12.7  
10:00 229.4 F 7.8  
12:00 240.6 F 13.0  
14:00 239.0 F 10.5  
16:00 237.0 F 8.7  
18:00 250.6 F 15.7  
20:00 262.7 F 22.6  
22:00 272.3 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.5 F 28.9  
2:00 271.5 F 28.4  
3:00 272.6 F 29.1  
4:00 271.7 F 28.3

Min 271.5 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.0 F 0.0  
5:00 215.5 F 0.0  
10:00 232.5 F 0.0  
15:00 236.1 F 0.0  
20:00 224.5 F 0.0  
25:00 220.3 F 0.0  
30:00 218.4 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 22:13  
STERILIZING: 4:00  
VENTING: 3:19  
DRYING: 30:00  
TOTAL CYCLE: 01:00:23  
CYCLE COMPLETE

06/13/2023 01:03 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_