

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000682  
7:13 22/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:25  
PRESSURIZING 3:25  
STERILIZING 4:27  
136.0 C 220kPa 4:27  
136.0 C 221kPa 4:57  
136.3 C 223kPa 5:27  
136.4 C 221kPa 5:57  
136.2 C 221kPa 6:27  
136.4 C 219kPa 6:57  
136.3 C 222kPa 7:27  
136.4 C 224kPa 7:57  
136.4 C 224kPa 7:57  
VENTING 7:57  
AIR DRYING 8:51  
CYCLE COMPLETE 9:14

-----  
Digital Signature #  
4D167DEA6EA5D83F  
-----