
Midmark M11 - v1.0.5
Total Cycles: 665

Name: RITTER2
Sterilizer ID: V2313228
12/29/2021 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.3 F 0.0
2:00 192.2 F 4.3
4:00 204.8 F 10.2
6:00 223.3 F 6.4
8:00 235.0 F 12.4
10:00 238.3 F 10.9
12:00 238.4 F 10.2
14:00 253.0 F 17.5
16:00 265.3 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 271.9 F 29.2
2:00 272.4 F 29.7
3:00 271.4 F 29.2
4:00 272.2 F 29.4

Min 271.3 F 28.9
Max 272.4 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.9 F 0.0
5:00 228.1 F 0.0
10:00 238.0 F 0.0
15:00 225.5 F 0.0
20:00 224.0 F 0.0
25:00 222.9 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 2:19
HEATING: 17:39
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:04
CYCLE COMPLETE

12/29/2021 02:35 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____