
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004705
13:30 14/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:18
STERILIZING 3:11
136.3 C 223kPa 3:11
136.3 C 224kPa 3:41
136.2 C 223kPa 4:11
136.4 C 225kPa 4:41
136.4 C 221kPa 5:11
135.9 C 217kPa 5:41
136.4 C 224kPa 6:11
136.4 C 224kPa 6:41
136.4 C 224kPa 6:41
VENTING 6:41
AIR DRYING 7:41
CYCLE COMPLETE 16:32

Digital Signature #
E3545AF28273411A
