
Midmark M11 - v1.0.5
Total Cycles: 2091

Name: RITTER2
Sterilizer ID: V2313228
02/27/2023 09:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.5 F 0.0
2:00 159.6 F 2.0
4:00 162.3 F 5.0
6:00 185.3 F 9.5
8:00 212.1 F 7.0
10:00 226.2 F 9.2
12:00 232.5 F 9.7
14:00 239.3 F 11.9
16:00 238.8 F 10.4
18:00 252.2 F 17.0
20:00 263.9 F 23.8
22:00 272.0 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.6 F 30.1
2:00 272.4 F 29.6
3:00 271.3 F 29.1
4:00 271.7 F 29.6

Min 271.0 F 29.0
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.4 F 0.0
5:00 238.9 F 0.0
10:00 226.8 F 0.0
15:00 212.5 F 0.0
20:00 203.3 F 0.0
25:00 193.6 F 0.0
30:00 196.3 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 22:10
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:23
CYCLE COMPLETE

02/27/2023 10:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____