
Midmark M11 - v1.0.5
Total Cycles: 2197

Name: RITTER2
Sterilizer ID: V2313228
03/30/2023 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.6 F 0.0
2:00 183.9 F 3.7
4:00 190.2 F 8.3
6:00 211.8 F 9.2
8:00 224.4 F 8.7
10:00 236.5 F 10.4
12:00 242.1 F 12.6
14:00 242.2 F 11.7
16:00 255.9 F 18.8
18:00 267.6 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.9
1:00 271.6 F 29.3
2:00 272.5 F 29.5
3:00 271.8 F 29.0
4:00 271.3 F 29.0

Min 271.2 F 28.7
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 217.2 F 0.0
10:00 231.3 F 0.0
15:00 226.3 F 0.0
20:00 227.6 F 0.0
25:00 229.8 F 0.0
30:00 225.3 F 0.0

Drying Complete

FILLING: 1:32
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:60
CYCLE COMPLETE

03/30/2023 10:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____