
Midmark M11 - v1.0.0
Total Cycles: 3645

Name: RITTER 1
Sterilizer ID: V2178198
08/14/2023 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.3 F 0.0
2:00 153.3 F 2.2
4:00 167.6 F 5.3
6:00 188.7 F 8.9
8:00 204.9 F 13.0
10:00 221.7 F 5.8
12:00 228.1 F 10.1
14:00 234.4 F 8.9
16:00 240.5 F 11.3
18:00 237.2 F 9.1
20:00 249.9 F 15.1
22:00 260.7 F 21.1
24:00 270.4 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 271.6 F 28.7
2:00 271.6 F 28.7
3:00 271.5 F 28.6
4:00 271.7 F 28.5

Min 271.5 F 28.5
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.4 F 0.0
5:00 227.0 F 0.0
10:00 216.9 F 0.0
15:00 205.5 F 0.0
20:00 190.6 F 0.0
25:00 214.8 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 25:28
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:04:22
CYCLE COMPLETE

08/14/2023 02:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____