

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1839

Name: RITTER2  
Sterilizer ID: V2313228  
12/14/2022 12:20 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.3 F 0.0  
2:00 190.8 F 4.0  
4:00 205.3 F 9.4  
6:00 220.4 F 7.1  
8:00 230.2 F 10.9  
10:00 232.9 F 8.5  
12:00 239.5 F 10.2  
14:00 247.8 F 14.6  
16:00 260.7 F 21.8  
18:00 272.1 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.0  
1:00 272.5 F 29.8  
2:00 271.9 F 28.9  
3:00 272.3 F 29.8  
4:00 272.2 F 29.3  
  
Min 271.5 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.6 F 0.0  
5:00 222.1 F 0.0  
10:00 242.7 F 0.0  
15:00 237.5 F 0.0  
20:00 226.6 F 0.0  
25:00 220.2 F 0.0  
30:00 217.8 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 18:23  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:28  
CYCLE COMPLETE

12/14/2022 01:17 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_