

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001059  
16:55 04/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:33  
STERILIZING 3:49  
136.0 C 222kPa 3:49  
136.4 C 225kPa 4:19  
136.4 C 225kPa 4:49  
136.4 C 225kPa 5:19  
136.3 C 224kPa 5:49  
136.2 C 224kPa 6:19  
136.3 C 225kPa 6:49  
136.3 C 224kPa 7:19  
136.3 C 224kPa 7:19  
VENTING 7:19  
AIR DRYING 8:13  
CYCLE COMPLETE 21:15  
-----

Digital Signature #  
CC00541EA9FF86C5  
-----