
Midmark M11 - v1.0.0
Total Cycles: 4153

Name: RITTER 1
Sterilizer ID: V2178198
12/22/2023 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.6 F 0.0
2:00 155.7 F 2.4
4:00 171.4 F 5.6
6:00 190.3 F 9.0
8:00 207.0 F 13.3
10:00 223.8 F 6.1
12:00 233.8 F 10.2
14:00 234.9 F 8.6
16:00 241.8 F 11.6
18:00 239.0 F 9.6
20:00 251.4 F 15.6
22:00 262.1 F 21.8
24:00 272.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.5 F 28.7
2:00 271.7 F 28.1
3:00 271.4 F 28.3
4:00 272.8 F 29.1

Min 271.0 F 27.8
Max 273.2 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.0 F 0.0
5:00 244.0 F 0.0
10:00 221.7 F 0.0
15:00 218.1 F 0.0
20:00 215.3 F 0.0
25:00 227.2 F 0.0
30:00 221.1 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 24:33
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:03:28
CYCLE COMPLETE

12/22/2023 12:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____