
Midmark M11 - v1.0.0
Total Cycles: 2086

Name: RITTER 1
Sterilizer ID: V2178198
07/05/2022 07:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.5 F 0.0
2:00 75.9 F 0.0
4:00 80.5 F 0.0
6:00 113.0 F 0.3
8:00 146.6 F 2.0
10:00 170.9 F 4.3
12:00 188.5 F 7.1
14:00 201.4 F 10.4
16:00 228.2 F 10.5
18:00 225.2 F 7.4
20:00 236.0 F 11.3
22:00 232.2 F 6.7
24:00 241.1 F 12.4
26:00 239.6 F 10.1
28:00 251.6 F 16.1
30:00 262.0 F 22.1
32:00 271.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 271.6 F 28.8
2:00 271.5 F 28.8
3:00 271.6 F 28.7
4:00 271.9 F 28.7

Min 271.5 F 28.7
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.9 F 0.0
5:00 230.6 F 0.0
10:00 228.5 F 0.0
15:00 223.5 F 0.0
20:00 198.7 F 0.0
25:00 202.7 F 0.0
30:00 192.4 F 0.0

Drying Complete

FILLING: 0:54
HEATING: 33:12
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:11:10
CYCLE COMPLETE

07/05/2022 08:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____