
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000363
15:00 20/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 2:14
STERILIZING 3:15
136.0 C 224kPa 3:15
136.4 C 233kPa 3:45
136.3 C 232kPa 4:15
136.4 C 233kPa 4:45
136.3 C 233kPa 5:15
136.0 C 231kPa 5:45
136.2 C 231kPa 6:15
136.4 C 233kPa 6:45
136.4 C 233kPa 6:45
VENTING 6:46
AIR DRYING 7:34
CYCLE COMPLETE 7:48

Digital Signature #
6091F7C5DFCFDE0C
