

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4706

Name: RITTER 1  
Sterilizer ID: V2178198  
07/09/2024 08:12 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.0 F 0.0  
2:00 69.6 F 0.2  
4:00 92.4 F 1.3  
6:00 123.6 F 3.4  
8:00 156.2 F 6.2  
10:00 182.7 F 9.4  
12:00 205.5 F 12.9  
14:00 217.2 F 3.6  
16:00 223.8 F 6.9  
18:00 232.6 F 10.4  
20:00 234.2 F 8.6  
22:00 238.2 F 10.0  
24:00 237.4 F 9.1  
26:00 245.1 F 12.7  
28:00 256.2 F 18.6  
30:00 266.3 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.1 F 28.4  
2:00 272.4 F 28.8  
3:00 271.5 F 28.6  
4:00 271.9 F 28.3

Min 271.2 F 28.2  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.6 F 0.0  
5:00 214.5 F 0.0  
10:00 212.9 F 0.0  
15:00 216.5 F 0.0  
20:00 218.2 F 0.0  
25:00 183.2 F 0.0  
30:00 175.1 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:40  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:09:53  
CYCLE COMPLETE

07/09/2024 09:21 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_