
Midmark M11 - v1.0.0
Total Cycles: 1077

Name: RITTER 1
Sterilizer ID: V2178198
08/09/2021 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.2 F 0.0
2:00 160.6 F 1.4
4:00 169.3 F 4.3
6:00 189.7 F 7.8
8:00 207.6 F 12.0
10:00 216.0 F 4.8
12:00 228.4 F 9.3
14:00 236.7 F 10.9
16:00 240.9 F 11.3
18:00 238.5 F 9.6
20:00 252.1 F 16.1
22:00 262.8 F 22.3
24:00 272.7 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.6 F 28.7
2:00 271.5 F 27.9
3:00 272.1 F 28.7
4:00 272.2 F 28.5

Min 271.3 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.2 F 0.0
5:00 210.9 F 0.0
10:00 202.4 F 0.0
15:00 194.6 F 0.0
20:00 191.3 F 0.0
25:00 193.0 F 0.0
30:00 198.9 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:21
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:02:44
CYCLE COMPLETE

08/09/2021 09:16 AM
ACCEPTED: X
REJECTED:
Operator: U01