
Midmark M11 - v1.0.5
Total Cycles: 2156

Name: RITTER2
Sterilizer ID: V2313228
03/16/2023 11:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.8 F 0.0
2:00 188.0 F 3.9
4:00 205.8 F 8.4
6:00 225.2 F 10.4
8:00 227.6 F 9.0
10:00 236.8 F 9.4
12:00 242.7 F 12.6
14:00 241.7 F 11.7
16:00 255.2 F 18.6
18:00 267.0 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 28.9
1:00 272.1 F 29.7
2:00 272.3 F 29.2
3:00 271.7 F 29.2
4:00 272.4 F 29.7

Min 271.5 F 28.8
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 223.7 F 0.0
10:00 220.2 F 0.0
15:00 209.5 F 0.0
20:00 209.5 F 0.0
25:00 220.5 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 19:19
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

03/16/2023 12:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____