

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
2.2uS / 1.4ppm  
CYCLE NUMBER 003691  
10:22 26/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:11  
PRESSURIZING 2:15  
STERILIZING 3:17  
136.0 C 220kPa 3:17  
136.0 C 223kPa 3:47  
136.4 C 225kPa 4:17  
136.3 C 223kPa 4:47  
136.4 C 224kPa 5:17  
136.3 C 223kPa 5:47  
136.4 C 223kPa 6:17  
136.4 C 224kPa 6:47  
136.4 C 224kPa 6:47  
VENTING 6:47  
AIR DRYING 7:38  
CYCLE COMPLETE 10:06

-----  
Digital Signature #  
7A66466C1B107EDC  
-----