
Midmark M11 - v1.0.5
Total Cycles: 3724

Name: RITTER2
Sterilizer ID: V2313228
06/20/2024 01:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 82.7 F 0.0
2:00 83.4 F 0.0
4:00 93.3 F 0.6
6:00 119.4 F 2.7
8:00 152.0 F 5.9
10:00 178.5 F 9.2
12:00 198.6 F 12.9
14:00 218.6 F 6.6
16:00 229.2 F 10.4
18:00 230.2 F 8.4
20:00 238.4 F 11.0
22:00 234.4 F 8.3
24:00 246.4 F 14.4
26:00 257.1 F 20.1
28:00 267.4 F 26.4
30:00 271.0 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.9
1:00 272.8 F 30.5
2:00 271.5 F 30.4
3:00 272.7 F 30.2
4:00 272.0 F 30.5

Min 271.0 F 29.9
Max 273.6 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.6 F 0.0
5:00 204.0 F 0.0
10:00 190.4 F 0.0
15:00 178.6 F 0.0
20:00 180.0 F 0.0
25:00 177.6 F 0.0
30:00 176.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:00
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:08:31
CYCLE COMPLETE

06/20/2024 03:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____