
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001531
13:11 10/25/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:24
PRESSURIZING 6:03
STERILIZING 7:32
136.0 C 220kPa 7:32
136.5 C 223kPa 8:02
136.5 C 223kPa 8:32
136.2 C 220kPa 9:02
136.4 C 222kPa 9:32
136.4 C 222kPa 10:02
136.4 C 223kPa 10:33
135.7 C 220kPa 11:02
VENTING 11:02
AIR DRYING 12:05
CYCLE COMPLETE 42:06

Digital Signature #
76FA429982480CA2
