

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3385

Name: RITTER2  
Sterilizer ID: V2313228  
02/14/2024 02:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 99.4 F 0.0  
2:00 101.4 F 0.0  
4:00 120.0 F 0.7  
6:00 155.2 F 3.4  
8:00 181.0 F 6.8  
10:00 198.0 F 10.5  
12:00 210.8 F 7.2  
14:00 218.9 F 7.1  
16:00 229.9 F 11.3  
18:00 230.5 F 7.8  
20:00 242.5 F 13.2  
22:00 242.1 F 11.8  
24:00 254.0 F 17.9  
26:00 264.7 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 273.0 F 29.8  
2:00 272.0 F 29.1  
3:00 271.9 F 29.5  
4:00 273.0 F 29.7

Min 271.5 F 28.9  
Max 273.1 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.5 F 0.0  
5:00 215.5 F 0.0  
10:00 222.7 F 0.0  
15:00 224.2 F 0.0  
20:00 223.3 F 0.0  
25:00 224.3 F 0.0  
30:00 224.5 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 27:59  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:06:27  
CYCLE COMPLETE

02/14/2024 03:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_