
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000663
12:24 20/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:37
PRESSURIZING 3:39
STERILIZING 4:44
136.5 C 223kPa 4:44
136.1 C 222kPa 5:14
136.3 C 225kPa 5:44
135.9 C 221kPa 6:14
136.4 C 221kPa 6:44
136.4 C 224kPa 7:14
136.3 C 222kPa 7:44
136.4 C 224kPa 8:14
136.4 C 224kPa 8:14
VENTING 8:15
AIR DRYING 9:12
CYCLE COMPLETE 9:13

Digital Signature #
C7733D32E9EC3597
