
Midmark M11 - v1.0.5
Total Cycles: 2404

Name: RITTER2
Sterilizer ID: V2313228
05/24/2023 07:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.7 F 0.0
2:00 79.3 F 0.0
4:00 93.1 F 0.0
6:00 113.6 F 0.0
8:00 155.7 F 2.0
10:00 177.7 F 5.4
12:00 196.3 F 9.2
14:00 212.9 F 12.9
16:00 221.4 F 6.4
18:00 231.6 F 10.6
20:00 232.8 F 8.1
22:00 243.2 F 12.7
24:00 241.4 F 11.2
26:00 253.7 F 17.4
28:00 264.4 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.3 F 28.9
2:00 272.2 F 29.3
3:00 271.7 F 28.5
4:00 272.5 F 29.3

Min 271.3 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.3 F 0.0
5:00 198.0 F 0.0
10:00 205.4 F 0.0
15:00 197.9 F 0.0
20:00 196.2 F 0.0
25:00 194.8 F 0.0
30:00 194.6 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 29:59
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:07:54
CYCLE COMPLETE

05/24/2023 08:41 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____