
Midmark M11 - v1.0.0
Total Cycles: 1646

Name: RITTER 1
Sterilizer ID: V2178198
02/17/2022 09:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.9 F 0.0
2:00 158.3 F 1.6
4:00 169.5 F 4.8
6:00 194.2 F 9.2
8:00 217.6 F 6.9
10:00 224.5 F 8.2
12:00 236.4 F 13.2
14:00 237.6 F 10.4
16:00 235.2 F 8.0
18:00 249.6 F 15.2
20:00 261.3 F 21.8
22:00 272.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.4 F 29.1
2:00 271.5 F 28.6
3:00 272.5 F 29.1
4:00 271.3 F 28.4

Min 271.3 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.9 F 0.0
5:00 217.9 F 0.0
10:00 228.0 F 0.0
15:00 236.7 F 0.0
20:00 218.0 F 0.0
25:00 216.8 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 22:30
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:48
CYCLE COMPLETE

02/17/2022 10:32 AM
ACCEPTED: X
REJECTED:
Operator: U02