
Midmark M11 - v1.0.5
Total Cycles: 4507

Name: RITTER2
Sterilizer ID: V2313228
04/11/2025 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.5 F 0.0
2:00 99.2 F 0.4
4:00 101.9 F 2.0
6:00 136.3 F 4.9
8:00 163.3 F 7.6
10:00 185.5 F 10.6
12:00 200.6 F 8.0
14:00 222.0 F 6.1
16:00 229.1 F 10.0
18:00 234.2 F 9.0
20:00 238.2 F 10.8
22:00 235.9 F 8.4
24:00 247.8 F 14.4
26:00 258.4 F 20.2
28:00 268.3 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.5
1:00 272.3 F 29.8
2:00 272.4 F 29.5
3:00 271.6 F 29.6
4:00 271.6 F 29.9

Min 271.5 F 29.4
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 231.8 F 0.0
10:00 223.7 F 0.0
15:00 232.9 F 0.0
20:00 222.5 F 0.0
25:00 218.3 F 0.0
30:00 216.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 29:55
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:08:51
CYCLE COMPLETE

04/11/2025 03:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____