
Midmark M11 - v1.0.0
Total Cycles: 2990

Name: RITTER 1
Sterilizer ID: V2178198
03/09/2023 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.3 F 0.0
2:00 159.9 F 3.0
4:00 181.0 F 6.9
6:00 204.8 F 11.7
8:00 218.9 F 6.0
10:00 233.3 F 11.2
12:00 232.9 F 7.8
14:00 242.4 F 11.6
16:00 245.0 F 12.7
18:00 257.9 F 19.6
20:00 269.4 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.3
1:00 272.5 F 29.1
2:00 271.5 F 28.2
3:00 272.6 F 29.1
4:00 271.7 F 28.3

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 227.7 F 0.0
10:00 235.3 F 0.0
15:00 226.5 F 0.0
20:00 226.8 F 0.0
25:00 227.2 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 20:51
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:60
CYCLE COMPLETE

03/09/2023 03:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____