
Midmark M11 - v1.0.0
Total Cycles: 5131

Name: RITTER 1
Sterilizer ID: V2178198
11/26/2024 10:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 69.2 F 0.0
4:00 86.3 F 0.0
6:00 121.2 F 0.0
8:00 147.3 F 0.1
10:00 180.2 F 0.4
12:00 194.4 F 0.8
14:00 208.7 F 3.5
16:00 219.4 F 6.3
18:00 226.8 F 9.6
20:00 236.6 F 13.6
22:00 236.6 F 9.8
24:00 238.7 F 9.6
26:00 245.9 F 13.3
28:00 244.4 F 12.4
30:00 256.9 F 18.9
32:00 268.0 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.2 F 28.5
2:00 272.4 F 28.9
3:00 271.6 F 28.1
4:00 272.4 F 28.7

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.5 F 0.0
5:00 207.7 F 0.0
10:00 205.6 F 0.0
15:00 211.1 F 0.0
20:00 208.6 F 0.0
25:00 205.3 F 0.0
30:00 202.7 F 0.0

Drying Complete

FILLING: 0:41
HEATING: 33:18
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:11:43
CYCLE COMPLETE

11/26/2024 11:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____