
Midmark M11 - v1.0.0
Total Cycles: 3477

Name: RITTER 1
Sterilizer ID: V2178198
07/04/2023 09:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.9 F 0.0
2:00 179.7 F 1.3
4:00 202.8 F 4.1
6:00 217.8 F 8.2
8:00 230.7 F 13.1
10:00 222.1 F 7.2
12:00 232.3 F 12.1
14:00 234.3 F 8.5
16:00 242.3 F 11.3
18:00 244.5 F 12.5
20:00 256.5 F 18.8
22:00 267.3 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.2 F 29.1
2:00 272.1 F 28.7
3:00 271.2 F 28.2
4:00 271.7 F 28.9

Min 271.1 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 236.8 F 0.0
10:00 230.3 F 0.0
15:00 228.4 F 0.0
20:00 228.8 F 0.0
25:00 227.6 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:31
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:42
CYCLE COMPLETE

07/04/2023 10:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____